



T&S BRASS AND BRONZE WORKS, INC.
 2 SADDLEBACK COVE / P.O. BOX 1088 / TRAVELERS REST, SC 29690
 PHONE 800-476-4103 FAX 864-834-3518



Model No.


BL-4010-01

Item No.:

Job Name:

Architect/Engineer Approval:

Notes:



NOTE:

1. MAXIMUM RECOMMENDED PRESSURE IS 125 PSI.
2. VALVE HAS A SPECIAL PACKING THAT IS SUITABLE FOR USE WITH OXYGEN, HYDROGEN, OR OTHER SPECIAL GASES.

Product Description:

LABORATORY NEEDLE HOSE COCK FOR USE WITH OXYGEN, HYDROGEN AND OTHER SPECIAL GASES

Drawn:

JKD

Approved

CA

Checked

MVW

Scale:

1' = 1'

Date:

4/22/98